



- NOTES:**
1. REFERENCE DATASHEET TA10601-1.
  2. ALL WIRING & CONNECTORS ARE SUPPLIED BY CUSTOMER AND MUST BE COMPLETED IN ACCORDANCE WITH APPLICABLE REGULATIONS.
  3. CAMERA PSU AND REGULATED LIGHT POWER SUPPLY ARE SUPPLIED IN IP20 RATED ENCLOSURES AND MUST BE MOUNTED IN A SUITABLE LOCATION.
  4. ETHERNET CAMERA IS SUPPLIED WITH RJ-45 JACK. CUSTOMER SUPPLIES CAT6 CABLING AS REQUIRED BETWEEN CAMERA & EXISTING ETHERNET NETWORK.
  5. FACTORY CALIBRATED WITY FORMAZIN STANDARDS.
  6. INFORMATION IS SUBJECT TO CHANGE WITHOUT NOTICE.

<p><b>NOTICE</b></p> <p>THIS DRAWING IS THE PROPERTY OF J.M. CANTY AND MUST BE RETURNED WITHOUT REPRODUCTION OR DUPLICATION, AT ANY TIME UPON REQUEST, BUT IN ANY EVENT AS SOON AS IT HAS SERVED THE PURPOSE FOR WHICH IT WAS FURNISHED, WHILE IN THE POSSESSION OF THE RECIPIENT, IT MUST BE PROPERLY SAFEGUARDED AGAINST DISCLOSURE TO ANYONE EXCEPT THOSE EMPLOYEES WHO REQUIRE IT FOR WORK OR JOB. THE RECIPIENT MUST KEEP CONFIDENTIAL AND REQUIRE HIS (HER) EMPLOYEES TO KEEP CONFIDENTIAL THE INFORMATION CONTAINED HERE ON.</p>	<p><b>IMPORTANT FOR THE REQUIRED MANUFACTURING SPECIFICATIONS SEE DOCUMENT #A4955</b></p> <p>NOTE: FAILURE TO RECEIVE DOCUMENT DOES NOT WAIVE VENDOR RESPONSIBILITY</p>	DRAWN BY	DATE	<p>J.M. CANTY INC. 6100 DONNER RD. BUFFALO, N.Y. 14294 U.S.A.</p>	<p><b>CANTY</b> WWW.JMCANTY.COM</p>	<p>J.M. CANTY INTL. LTD. BALLYCOOLIN BUS. PK. BLANCHARDSTOWN DUBLIN 15, IRELAND</p>
		APPROVAL	DATE			
		DATE	DATE			
		ENGINEERING ANALYSIS	DATE	TITLE		
		CURRENT	DATE	TYPICAL IN-LINE TURBIDITY ANALYZER		
		PENDING	DATE	SALES/DATA SHEET, FLANGE CONNECTION		
		QUALITY CONTROL	DATE	SIZE (REVISION)	FILE NAME	
		I.O.A.	DATE	D 0	VD10977-300	
				SCALE AS NOTED		SHEET 1 OF 1